

कार्यालय, प्रधानाचार्य एवं मुख्य कार्यपालन अधिकारी
हकीम सैयद जियाउल हसन, शासकीय (स्वशासी) यूनानी चिकित्सा महाविद्यालय
(कोलार-नेहरू नगर बायपास रोड, मेनित के पीछे, भोपाल 462003)

Web: www.gumcbhopal.in

E.mail: gumcbhopal@gmail.com

दूरभाष क्र. -0755-2970354

(महाविद्यालय हेतु 100 एम.ए. की डिजिटल एक्स रे मशीन के कय हेतु ई-निविदा के नियम एवं शर्तें)

हकीम सैयद जियाउल हसन शासकीय (स्वशासी) यूनानी चिकित्सा महाविद्यालय भोपाल में 100 एम.ए. की डिजिटल एक्स रे मशीन के कय हेतु ई-निविदा प्रक्रिया के अंतर्गत निविदाएं आमंत्रित की जाती हैं, जो कि <https://mptenders.gov.in> पर निर्धारित समयावधि में उपलब्ध हैं। निविदा प्रस्तुत करने के नियम निर्देश एवं शर्तें महाविद्यालय की वेबसाइट www.gumcbhopal.in पर भी प्रदर्शित हैं।

1. ई-निविदा प्रपत्र प्राप्त करने एवं जमा करने की अंतिम तिथि ~~30-09-2025~~ ^{14/11/2025} को सायं 05.00 बजे तक है एवं ई निविदा खोले जाने की तिथि दिनांक ~~01-10-2025~~ ^{16/11/2025} दोपहर 02.00 बजे है।
2. ई-निविदा बिड दस्तावेज हस्तांतरणीय नहीं हैं। ~~31-10-2025~~ ^{16/11/2025}
3. इच्छुक ई-निविदाकर्ता ऑनलाइन mptenders.gov.in से ई-निविदा राशि रु. 2500/- का भुगतान कर सकते हैं, जिसको mptenders.gov.in पर तकनीकी एवं वित्तीय निविदा पृथक-पृथक ऑनलाईन सबमिट की जाना है।
4. ई.एम.डी.राशि रुपये 20000.00 शब्दों में (बीस हजार मात्र) जिसे ई-पेमेन्ट के माध्यम से mptenders.gov.in पर tender सबमिट करते समय पेमेन्ट करना होगा। ई.एम.डी.राशि सफल निविदाकार को अनुबंध होने पर नियमानुसार वापस की जावेगी, शेष निविदाकारों की ई.एम.डी.राशि mptenders.gov.in के माध्यम से लौटा दी जावेगी। सफल निविदाकार को निष्पादन प्रतिभूति राशि (अनुबंध राशि के 3 प्रतिशत राशि रुपये 30000.00 तक) बैंक ड्राफ्ट के माध्यम से महाविद्यालय में जमा करना होगी। निष्पादन प्रतिपूर्ति राशि बिना ब्याज के अनुबंध समाप्त होने पर वापस लौटा दी जावेगी। म.प्र.के लघु उद्यम तथा स्टार्टअप को ई.एम.डी. राशि के भुगतान में छूट रहेगी। ई.एम.डी. राशि छूट प्राप्त करने हेतु आवश्यक दस्तावेज ई-निविदा में अपलोड करना अनिवार्य है।
5. अनुबंधित फर्म को उचित गुणवत्ता की सामग्री की आपूर्ति अनिवार्य है। इस हेतु कोई भाड़ा/हम्माली पृथक से देय नहीं होगी।
6. वित्तीय निविदा प्रपत्र-02 अनुसार दरें <https://mptenders.gov.in> पर ई-निविदा की एक्सल शीट BOQ में ही अंकित करना है।
7. सामग्री की गुणवत्ता संबंधी प्रमाणीकरण मान्य होने के पश्चात् ही भुगतान की कार्यवाही की जावेगी, जी.एस.टी.नियमानुसार देय होगी।
8. सामग्री आपूर्ति को व्यवहारिक एवं आसान किये जाने हेतु ई वित्तीय निविदा का प्रपत्र -02 अनुसार सामग्री की दरों में सकल (Aggregate) दर अनुसार एल-1 को निर्धारित किया जा सकेगा।
9. किसी भी प्रकार की शिकायत प्राप्त होने पर तथा नियम शर्तों का पालन न करने पर अथवा सन्तोषजनक सेवा न पाये जाने पर अनुबंध तत्काल निरस्त कर दिये जावेंगे एवं सम्पूर्ण जिम्मेदारी संबंधित फर्म की होगी। ऐसी स्थिति में जमा निष्पादन प्रतिभूति की राशि राजसात कर ली जावेगी, तथा संबंधित फर्म पर ब्लैक लिस्टिंग की कार्यवाही की जावेगी, एवं सामग्री की आपूर्ति हेतु एल-02 फर्म से अनुबंध किया जायेगा।
10. सफल निविदाकार से किया गया अनुबंध आगामी वर्ष तक की अवधि तक के लिये मान्य होगा।

- स्वीकार होने पर सफल निविदाकार को एक अनुबंध राशि रुपये 100.00 के नॉन ज्युडीशियल स्टाम्प पेपर पर नोटरी कर प्रस्तुत करना अनिवार्य है। जिसका व्यय निविदाकार को करने करना होगा।
12. निविदाकर्ता फर्म भोपाल नगर निगम सीमा में स्थापित होना चाहिये जिसका गुमास्ता तकनीकी निविदा के अभिलेखों के साथ संलग्न करना आवश्यक होगा।
 13. निविदाकर्ता के पास स्वयं अथवा फर्म का पैन नंबर होना चाहिये जिसकी प्रतिलिपि तकनीकी निविदा के अभिलेखों के साथ संलग्न करना आवश्यक होगा।
 14. निविदाकर्ता के पास जी.एस.टी. नम्बर होना आवश्यक है जिसकी प्रतिलिपि तकनीकी निविदा के अभिलेखों के साथ संलग्न करना आवश्यक होगा।
 15. निविदाकर्ता फर्म का वित्तीय वर्ष 2024-25 का टर्न ओवर न्यूनतम 20 लाख होना आवश्यक है, जिसके साक्ष्य के रूप में वित्तीय वर्ष 2024-25 की आयकर विवरणी की प्रति, बैलेंस शीट, लाभ-हानि पत्रक/चाटेर्ड एकाउन्टेन्ट द्वारा प्रमाणित एवं वित्तीय वर्ष 2024-25 की आयकर विवरणी तकनीकी निविदा के दस्तावेजों के साथ अपलोड करना आवश्यक है।
 16. निविदाकर्ता द्वारा कार्यादेश का कुछ भाग या पूर्ण भाग अन्य किसी ठेकेदार/व्यक्ति को नहीं सौंपेगा यदि किसी निविदाकर्ता द्वारा किसी अन्य फर्म/व्यक्ति को कार्य सौंपा जाता है, या कार्य कराता पाया जाता है तो उसका अनुबंध समाप्त कर फर्म को Black listed किया जायेगा।
 17. आवेदनकर्ता फर्म प्रदेश के किसी भी शासकीय संस्था से ब्लैक लिस्टेड नहीं होनी चाहिए। यदि जानकारी प्राप्त होती है, तो निष्पादन प्रतिभूति राशि राजसात कर संबंधित फर्म की निविदा/अनुबंध निरस्त कर दिया जावेगा। ब्लैक लिस्टेड न होने का घोषणा/शपथ पत्र राशि रु. 100.00 एक सौ मात्र के नॉन ज्युडीशियल स्टाम्प पर नोटराइज कराकर तकनीकी निविदा के साथ संलग्न करना अनिवार्य होगा।
 18. ई-निविदा खोलते समय निविदाकर्ता अथवा उनके प्रतिनिधि (अधिकृत पत्र प्रस्तुत करने पर) शामिल हो सकते हैं।
 19. ई-निविदा को स्वीकृत अथवा अस्वीकृत करने का सम्पूर्ण अधिकार प्रधानाचार्य को होगा, जो सभी को मान्य एवं बंधनकारी होगा।
 20. किसी भी विवाद की स्थिति में प्रधानाचार्य का निर्णय अंतिम एवं सर्वमान्य होगा, इसका पालन करना होगा।
 21. ई-तकनीकी निविदा में जो दस्तावेज अपलोड किये हैं, उन दस्तावेजों की सत्यापित प्रति का 01 सेट ई-तकनीकी निविदा खोलते समय उपलब्ध कराना अनिवार्य होगा।
 22. अन्य कोई शर्त आवश्यकता पड़ने पर मान्य करना होगी।
 23. ई-तकनीकी निविदा का विश्लेषण करते समय केवल वांछित दस्तावेजों (चैक लिस्ट के अनुसार) को ही विचार क्षेत्र में लिया जावेगा।
 24. किसी भी विवाद की स्थिति में न्याय क्षेत्र भोपाल रहेगा।

प्रधानाचार्य एवं मुख्य कार्यपालन अधिकारी
हकीम सैयद जियाउल हसन
शासकीय(स्वशासी) यूनानी चिकित्सा
महाविद्यालय, भोपाल।



(ई-वित्तीय निविदा का प्रपत्र-02)

X Ray Machine Rate List

S.No	Item Title	Rate
1	X Ray Machine Rates as per Technical Specification	

नोट :-उक्त प्रपत्र प्रारूप मात्र है इस प्रपत्र के अनुसार दर राशि को ई-वित्तीय निविदा ऑनलाइन www.mptender.gov.in पर तकनीकी निविदा की एक्सल शीट BOQ में ही दर्ज करना है।

(डॉ० मेहमूदा बेगम)
प्र०प्रधानाचार्य / मुख्य कार्यपालन अधिकारी
हकीम सैयद जियाउल हसन शासकीय / स्वशासी
यूनानी चिकित्सा महाविद्यालय, भोपाल

  

(ई-वित्तीय निविदा का प्रपत्र-02)

X Ray Machine Rate List

S.No	Item Title	Rate
1	X Ray Machine Rates as per Technical Specification	

नोट :-उक्त प्रपत्र प्रारूप मात्र है इस प्रपत्र के अनुसार दर राशि को ई-वित्तीय निविदा ऑनलाइन www.mptender.gov.in पर तकनीकी निविदा की एक्सल शीट BOQ में ही दर्ज करना है।

(डॉ० मेहमूदा बेगम)

प्र०प्रधानाचार्य / मुख्य कार्यपालन अधिकारी
हकीम संयद जियाउल हसन शासकीय /स्वशासी
यूनानी चिकित्सा महाविद्यालय, भोपाल

  

Below is a **technical specification draft** for the **1417 DR Panel**, suitable for inclusion in a government procurement bid includes key technical, operational, and environmental details.

1417 Flat-Panel Digital X-Ray Detector

For Government Tender-Technical Specification

1. General Description

- **Device Type:** Flat-panel digital X-ray detector (DR panel); wireless/wired configuration; suitable for portable, rugged, and harsh environments
- **Series:** Designed with a non- glass TFT substrate (a-Si or IGZO) for enhanced durability and mobility

2. Certifications: Registered under FDA 510(k)(K211108).

3. Imaging & Sensor Specifications

Parameter	Description
Scintillator	CesiumIodide (CsI) or Gadox(GOS),high X-ray sensitivity and image quality.
Sensor Array	a-Si TFT or IGZO non-glass substrate
Image Size	14"× 17"(approx.350× 427mm)
Resolution	2,560×3,072 pixels(≈7.8 Mpx)
Pixe lPitch	140μm
Spatial	3.6 lp/mm(approx.)
Resolution	14-bit or 16-bit format
Image Data Depth	



4. Triggering, Calibration & Performance

- **Trigger Modes:** Auto Exposure Detection (AED), Manual Trigger.
- **Calibration Modes:** Automatic Calibration Control (ACC) and manual calibration supported.
- **Capture Cycle Time:** Approximately 3–5 seconds per image.
- **Data Transfer Time:** <1 second over wired interface.

5. Connectivity & Interfaces

- **Wired Interface:** Gigabit Ethernet via HDMI-type data cable; AC power 100–240V (50/60 Hz) or DC 18 V/3.5 A.
- **Wireless Interface:** IEEE802.11ac Wi-Fi (for high-speed wireless operation), suitable for mobile and stand-alone use.

6. Storage & Viewing

- **On-Board Storage:** Can store 150–200 images on board; enables stand-alone functioning without workstation.
- **Web Viewer:** Built-in capability to view images in near real-time using mobile devices (iOS, Android, Windows) via web-based viewer.

7. Design Battery & Durability

- **Housing:** One-piece carbon fiber or magnesium composite body; light weight yet durable; supports up to 500 kg load (whole surface), 180 kg on 4 × 4 cm center.
- **IP Rating:** IP67 dust and water proof rating—resists cleaning with liquid sand harsh environments.
- **Battery & Power:** Lithium-ion polymer battery; lasts up to ~6 hours per battery (wireless Wi-Fi mode). Optional dual batteries extend operation time up to ~14 hours; can charge inside detector using dual battery charger.
- **Battery Charger:** OLED display status, 16.78 V/1.5 A charging; 4 h charging time.
- **Carry Ergonomics:** Easy-grip rounded edges and removable handle for secure handling.

8. Environmental & Physical Specs

- **Operation Temp:** Typically 10 °C to 35 °C; humidity 20%–75% RH (non-condensing).
- **Storage Conditions:** –10 °C to 60 °C; up to 90% RH (non-condensing).
- **Dimensions:** 385 mm × 460 mm × 15 mm (W × L × T).
- **Weight:** Approx. 2.4 kg (without battery), up to 2.8 kg including battery.

9. Regulatory & Safety Compliance

- **CE/Medical Directive:** Meets EU Class II requirements under 93/42/EEC directive.
- **EMC/RF:** FCC Part 15 compliant (RF exposure statements included in user manual).



- **FDA 510(k):** Substantial equivalence confirmed under K211108

Summary Highlights for Bid Submission

- **Highly Mobile & Rugged:** Non-glass TFT sensor, light weight carbon/magnesium body, IP67 rating make it ideal for field deployment or mobile radiography units.
- **Flexible Power & Connectivity:** Both wired (Gigabit Ethernet) and wireless (802.11ac Wi-Fi) modes supported; battery allows extended use up to 14hrs.
- **User-Friendly Operation:** On-board image storage and web-viewer support stand-alone functionality—no console needed. Ergonomic design simplifies handling.
- **Fast & High-Quality Imaging:** Advanced Gadox / CsI scintillator, AED trigger, sub-5 sec capture time, with calibrated accuracy and low noise at low dose.
- **Compliance-Ready:** Certified to CE Class IIa, FCC, and FDA 510(k) standards—ready for regulatory acceptance.

1) General

- Device type: Wireless flat-panel digital X-ray detector for general/dynamic radiography, cassette size 14×17 in. Supports auto exposure detection (AED) and manual trigger modes.
- Detector technology: IGZO or a-Si TFT array; conversion screen CsI or GOS (Gadox).

- Image data depth: 16-bit.

2) Imaging Performance

- Active area (detector size): 14 × 17 in (approx. 350 × 427 mm).
- Sensor pixels / matrix: Supports 3534 × 4302 and 2500 × 3052 modes.
- Capture cycle: 2–5 seconds (approx.).
- Calibration modes: Automatic Calibration Control (ACC) and manual calibration.

3) Connectivity & Work-flow

- Wired interface: Gigabit Ethernet
- Wireless workflow: Wireless operation using embedded AP with bridge/router or direct laptop connection ; IEEE 802.11ac recommended.

4) Power & Battery

- Battery type/capacity: Lithium-ion polymer battery, 14.8 Vdc, 3300 mAh.
- Battery life: Approx. 8 hours of continuous operation under standard use.
- External adaptor (for PICB/tether): 18 Vdc, 3.5 A (Sin pro HPU63A-107 or equivalent).

5) Physical & Environmental

- Dimensions (W × L × T): 382 × 460 × 15 mm (cassette size).
- Weight: 3.3 kg ±10% (with battery).
- Operating conditions: Temperature 10–35 °C; Humidity 20–75% RH; Pressure 700–1060 hPa.
- Transport/storage conditions: Temperature –20–70 °C; Humidity 10–95% RH; Pressure 400–1500 hPa.



6) Safety & Regulatory

- Complies with medical electrical safety and EMC requirements.
- Tested as a Class A digital device per FCC Part 15.
- RF exposure compliance with FCC/IC guidelines.

7) Supplied/Compatible Accessories

- PICB (Interface Box) for tethered use: 204 × 112 × 36 mm; weight ~300 g.
- Adaptor: HPU63A-107, input 100–240 Vac 47–63 Hz 1.62–0.72 A; output 18 V / 3.5 A.

Features that Differentiate 1417

- IGZO TFT option with higher sensitivity and SNR for low-dose imaging.
- Non-glass TFT substrate option reduces breakage risk in portable/mobile use.
- Embedded AP & flexible wireless topologies for robust portable/mobile workflows.
- Fast, stable AED with multiple trigger modes (Auto, Continuous Auto, Manual, Continuous Manual).
- Lightweight, slim cassette form factor (~3.3 kg, 15 mm thick) with handle/grips.

Bid Compliance Language

- The offered detector shall be 1417 or a device that meets or exceeds all specifications listed above, without exception.
- If 'or equal' is proposed, the vendor must document identical or superior values for detector technology, active area, pixel matrix, capture cycle, bit depth, battery life, dimensions/weight, workflows, and environmental limits.
- Vendor must include data sheets/manuals and provide 5 year of warranty with 5 year of insurance for the DR panel and replacement or repair and maintained whichever required on site and local service details.
- Vendor should provide DR panel protection cover & led apron to protect the person & panel from damages.

Three handwritten signatures in blue ink are visible. The first signature on the left is a stylized, somewhat illegible mark. The middle signature is more legible, appearing to be 'Liam'. The third signature on the right is also stylized and illegible.

Technical Specification— 1417DR Panel(14×17in)

1) General

- **Device type:** Wireless flat-panel digital X-ray detector for general/dynamic radiography , cassette size 14×17in. Must support auto exposure detection(AED) and manual trigger modes.
- **Detector technology:** IGZO or a-Si TFT array; conversion screen CsI or GOS (Gadox).
- **Image data depth:** 16-bit.

2) Imaging Performance

- **Active area (detector size):** 14 × 17 in (approx. 350 × 427 mm).
- **Sensor pixels/matrix:** Must support the 1417 pixel formats 3534×4302 and 2500 × 3052 (per factory modes).
- **Capture cycle:** 2–5 seconds (approx.).
- **Calibration modes:** Automatic Calibration Control (ACC) and manual calibration.

3) Connectivity & Workflow

- **Wired interface:** Gigabit Ethernet (for tethered operation via PICB interface box).
- **Wireless work flow:** Must support wireless operation using the detector's embedded AP with bridge/router or direct connection to a laptop for portable/mobile X-ray systems; IEEE 802.11ac adapter recommended for the workstation.
- **Acquisition utilities:** Must be compatible with probe for connection setup, testing, and calibration file creation.

4) Power & Battery

- **Battery type/capacity:** Lithium-ion polymer battery, 14.8Vdc, 3300mAh (per pack).
- **Battery life :** Approx. 8 hours of continuous operation under standard use.



- **External adaptor (for PICB/tether):** 18Vdc,3.5A(Sinpro HPU63A-107or equivalent).

5) Physical & Environmental

- **Dimensions (W×L×T):** 382×460×15mm(panel body; cassette size), with handle/grips as illustrated in PIXXGEN dimensional drawings.
- **Weight:**3.3kg ±10%(with battery).
- **Operating conditions:**
 - Temperature:10–35°C
 - Relative humidity: 20–75%RH
 - Atmospheric pressure:700–1060hPa
- **Transport/storage conditions:**
 - Temperature:–20–70°C
 - Relative humidity: 10–95%RH
 - Atmospheric pressure:400–1500hPa

6) Safety & Regulatory (Detector & Wireless)

- Must comply with applicable medical electrical safety and EMC requirements referenced device test data Class A digital device per FCC Part 15.
- **RF exposure & channelization:** Operation conforms to FCC/IC RF exposure guidelines; U.S./Canada operation limited to channels1–11;5.15–5.25GHzrestricted to indoor use when applicable.

7) Supplied/Compatible Accessories

- **PICB(for tethered use:**
 - Dimensions:~204 ×112 ×36 mm;weight~300g
 - Adaptor:HPU63A-107,input 100–240Vac47–63Hz1.62–0.72A;output 18 V / 3.5 A.

Features that Differentiate 1417(helpful for justification)

- **IGZO TFT option with high sensitivity and SNR:** series offers IGZO TFT (in addition to a-Si), improving SNR/DQE versus conventional a-Si, which is valuable for low-dose imaging.



- **Non-glass TFT substrate option:** literature highlights non-glass substrate construction in the 1417 series, designed to reduce breakage risk in mobile/portable use.
- **Embedded AP & flexible wireless topologies:** Detector supports embedded-AP workflows (bridge/station modes) and direct laptop connection, enabling robust portable or mobile deployments without generator cabling.
- **Fast, stable AED with continuous auto mode:** Multiple trigger modes (Auto, Continuous Auto Manual, Continuous Manual) support reliable acquisition across varied generators and use cases.
- **Lightweight, slim cassette form factor:** ~3.3kg and 15mm thick with handle/grips for safer handling in emergency and bedside scenarios.

Bid Compliance

1. The offered detector **shall be 1417** or a device that **meets or exceeds all** specifications listed in Sections 1–7 above, **without exception**.
2. If “or equal” is proposed, the vendor **must document** identical or superior values for: detector technology (IGZO/a-Si), active area, pixel matrix, capture cycle, 16-bit depth, battery life, dimensions/weight, wired/wireless workflows (incl. embedded-AP operation), and environmental limits.
3. Vendor must include manufacturer’s data sheets/manual pages substantiating all claimed specifications and provide warranty and local service details.



// शपथ -पत्र //

मैं पिता श्री.....
 निवासी
 उम्र वर्ष शपथ पूर्वक कथन करता हूं कि हकीम सैयद ज़ियाउल हसन शासकीय (स्वशासी)यूनानी चिकित्सालय में एक्स रे मशीन के स्थानीय क्रय हेतु निविदा में दी गई समस्त शर्तों का पूर्ण होशो-हवास में अध्ययन कर लिया है जिसकी समस्त शर्तें मुझे मान्य हैं। एवं मेरे द्वारा प्रस्तुत समस्त दस्तावेज वैधानिक हैं एवं किसी भी शासकीय संस्था द्वारा मेरी फर्म को ब्लैक लिस्टेड नहीं किया गया है। निविदा की शर्तों का पालन नहीं करने एवं मेरे द्वारा प्रस्तुत दस्तावेज त्रुटिपूर्ण पाये जाने की स्थिति में यदि मेरी निविदा/अनुबंध निरस्त की जाती है, तो मुझे किसी प्रकार की आपत्ति नहीं होगी।

गवाह/साक्षी के हस्ताक्षर
 गवाह/साक्षी का नाम
 मोबाइल नं०

निविदाकर्ता के हस्ताक्षर
 निविदाकर्ता का नाम
 मोबाइल नं०

नोट :- उक्त शपथ पत्र राशि रु 100.00 एक सौ मात्र के नॉन ज्युडीशियल स्टॉप पर नोटराइज कराकर तकनीकी निविदा के साथ संलग्न करना अनिवार्य होगा ।

// चैक-लिस्ट //

उपरोक्त वर्णित नियम शर्तों अनुसार निम्न दस्तावेजों को तकनीकी निविदा प्रपत्र 01 अनुसार सबमिट करते समय आवश्यक रूप से संलग्न किया जाना है :-

क्र	दस्तावेज का नाम	संलग्न हाँ/नहीं
01	जीवित गुमास्ता संबंधित कार्य का होना आवश्यक है जिसकी छाया प्रति	
02	पैन कार्ड की छाया प्रति	
03	जी.एस.टी.पंजीयन पत्र की प्रति (जी.एस.टी.जमा सम्बंधी विगत 02 वित्तीय वर्षों का ऑनलाईन पत्रक तकनीकी निविदा के दस्तावेजों के साथ संलग्न करना अनिवार्य है)	
04	निविदाकर्ता फर्म का वित्तीय वर्ष 2024-25 का टर्न ओवर न्यूनतम 20 लाख होना आवश्यक है जिसके साक्ष्य के रूप में विगत तीन वित्तीय वर्षों की आयकर विवरणी, लाभ-हानि पत्रक चार्टर्ड एकाउंटेंट द्वारा प्रमाणित	
05	ई.एम.डी.राशि की ई-प्रिंटेड प्रति (प्रधानाचार्य हकीम सैयद जियाउल हसन शासकीय/स्वशासी यूनानी चिकित्सा महाविद्यालय भोपाल के नाम से) अथवा एम.एस.एम.ई.पंजीयन प्रमाण पत्र	
06	निविदाकार फर्म किसी शासकीय/अशासकीय/अर्धशासकीय संस्था द्वारा ब्लैक लिस्ट नही होना चाहिए इस हेतु शपथ पत्र की प्रति प्रपत्र - 03 अनुसार	
07	कंपनी का प्राधिकरण (Authorization) संलग्न होना चाहिए।	

निविदाकर्ता के हस्ताक्षर

निविदाकर्ता का नाम

मोबाइल नं०

नोट :-चैक लिस्ट के सरल क्र-01 से 07 तक के अभिलेखों के आधार पर ही तकनीकी निविदा का विश्लेषण किया जावेगा इन्ही अभिलेखों के आधार पर तकनीकी निविदा मान्य/अमान्य की जायेगी इनके अतिरिक्त किन्ही अन्य अभिलेखों को विचार क्षेत्र में नहीं लिया जावेगा ।

